

Print this Page

**CHEST RESEARCH FOUNDATION
ONLINE SPIROGRAM QUALITY ASSESSMENT SYSTEM**



Report Code : 36

Spirogram Code : 80/10/PC79006

Trial Code : CP/01/08

Approval Status : Not Approved

Chest Research Foundation Assessor Code : Dr. Salvi

- Reasons For Not Approval:
- Inadequate inhalation
 - Poor effort
 - Delayed start: Excessive hesitation, air leak
 - Coughing: in the first second or any other cough, which according to technician will, interfere with the measurements

Disclaimer:

- * This spirogram quality assessment system does not accept responsibility for incorrect information related to trial code, center code and investigator code entered by the user.
- * The users of this system to note that spirogram quality assessment system is purely made for acceptability.

[Close](#)